

SPONSOR 10991



3D Systems Corporation  
333 Three D Systems Circle  
Rock Hill, SC 29730  
United States

**REPORT DATE**

November 11, 2016

Amended, rev. 2

**CERTIFICATE**  
**CLASS VI - 37°C PLASTIC TESTS**  
**(USP 34 <88>, REV. 5/2011)**

Name: Accura® ClearVue  
Physical Description: Solid  
Lot Number: 160829-2  
Storage Conditions: Room Temperature  
Date Received: June 27, 2011

Systemic Injection Test in CD-1 Mice  
Pacific BioLabs Report No.: 11F0367H-X02G

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits  
Pacific BioLabs Report No.: 11F0367H-X03G

Implantation Test in New Zealand White Rabbits  
Pacific BioLabs Report No.: 11F0367H-X04G

The test article met the requirements for Plastic Class VI - 37°C testing as per USP <88> guidelines.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist

11/11/16

X

I approve the content of this document.  
Signed by: Zuzana Karjala

11/11/16

X

QA Review  
Signed by: Kenneth Tan

:ahc