

SPONSOR 10991



3D Systems Corporation  
26600 SW Parkway Ave.  
Suite 300  
Wilsonville, OR 97070  
United States

REPORT DATE

October 05, 2017

**CERTIFICATE**  
**CLASS VI – 50°C PLASTIC TESTS**  
**(USP 40 <88>, REV. 5/2017)**

Test Article Name: CR-CL 200, BW170907  
Physical Description: Solid  
Lot Number: LC172505W  
Sample Code: Not provided by Sponsor  
Part Number: Not provided by Sponsor  
Expiration Date: Sept. 6<sup>th</sup>, 2020  
Sterilization Data: Non-sterile  
Storage Conditions: Room Temperature  
Date Received: September 08, 2017

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits  
Pacific BioLabs Report No.: 17I0109H-X02G

Systemic Injection Test in CD-1 Mice  
Pacific BioLabs Report No.: 17I0109H-X03G

Implantation Test in New Zealand White Rabbits  
Pacific BioLabs Report No.: 17I0109H-X04G

The test article met the requirements for Plastic Class VI – 50°C testing as per USP <88> guidelines.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist  
10/06/17

X

I approve the content of this document.  
Signed by: Zuzana Karjala

10/05/17

X

QA Review  
Signed by: Sarah Lauder

:af