

SPONSOR 10991



3D Systems Corporation
26600 SW Parkway Ave.
Suite 300
Wilsonville, OR 97070
United States

REPORT DATE

October 05, 2017

CERTIFICATE
CLASS VI – 50°C PLASTIC TESTS
(USP 40 <88>, REV. 5/2017)

Test Article Name: CR-WT 200, BW170901
Physical Description: Solid
Lot Number: LW170507W
Sample Code: Not provided by Sponsor
Part Number: Not provided by Sponsor
Expiration Date: Aug. 31st, 2020
Sterilization Data: Non-sterile
Storage Conditions: Room Temperature
Date Received: September 05, 2017

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 17I0034H-X02G

Systemic Injection Test in CD-1 Mice
Pacific BioLabs Report No.: 17I0034H-X03G

Implantation Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 17I0034H-X04G

The test article met the requirements for Plastic Class VI – 50°C testing as per USP <88> guidelines.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist
10/06/17

X

I approve the content of this document.
Signed by: Zuzana Karjala

10/05/17

X

QA Review
Signed by: Sarah Lauder

:af