

SPONSOR 10991



3D Systems Corporation
333 Three D Systems Circle
Rock Hill, SC 29730
United States

REPORT DATE

November 11, 2016

Amended, rev. 3

CERTIFICATE
CLASS VI - 37°C PLASTIC TESTS
(USP 34 <88>, REV. 5/2011)

Name: VisiJet® Clear
Physical Description: Solid
Lot Number: 160829-2
Storage Conditions: Room Temperature
Date Received: June 27, 2011

Systemic Injection Test in CD-1 Mice
Pacific BioLabs Report No.: 11F0367H-X02G

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 11F0367H-X03G

Implantation Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 11F0367H-X04G

The test article met the requirements for Plastic Class VI - 37°C testing as per USP <88> guidelines.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLatg, Senior Scientist

11/11/16

X

I approve the content of this document.
Signed by: Zuzana Karjala

11/11/16

X

QA Review
Signed by: Kenneth Tan

:ahc