

SPONSOR 10991



3D Systems Corporation
26600 SW Parkway Ave.
Suite 300
Wilsonville, OR 97070
United States

REPORT DATE

July 26, 2017

CERTIFICATE
CLASS VI - 50°C PLASTIC TESTS
(USP 40 <88>, REV. 5/2017)

Test Article Name: M2R-GRY BW170620
Physical Description: Solid
Lot Number: RR172505W
Sample Code: Not provided by Sponsor
Part Number: Not provided by Sponsor
Expiration Date: June 19th, 2020
Sterilization Data: Non-sterile
Storage Conditions: Room Temperature
Date Received: June 21, 2017

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 17F0386H-X02G

Systemic Injection Test in CD-1 Mice
Pacific BioLabs Report No.: 17F0386H-X03G

Implantation Test in New Zealand White Rabbits
Pacific BioLabs Report No.: 17F0386H-X04G

The test article met the requirements for Plastic Class VI – 50°C testing as per USP <88> guidelines.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist

07/27/17

X

I approve the content of this document.
Signed by: Zuzana Karjala

07/26/17

X

QA Review
Signed by: Vicky Tai
:af