

SPONSOR 10991



3D Systems Corporation  
26600 SW Parkway Ave.  
Suite 300  
Wilsonville, OR 97070  
United States

**REPORT DATE**

August 25, 2017

**CERTIFICATE**  
**CLASS VI – 50°C PLASTIC TESTS**  
**(USP 40 <88>, REV. 5/2017)**

Test Article Name: M2R-TN BW170711  
Physical Description: Solid  
Lot Number: DY17202A  
Sample Code: Not provided by Sponsor  
Part Number: Not provided by Sponsor  
Expiration Date: July 10, 2020  
Sterilization Data: Non-sterile  
Storage Conditions: Room Temperature  
Date Received: July 12, 2017

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits  
Pacific BioLabs Report No.: 17G0155H-X02G

Systemic Injection Test in CD-1 Mice  
Pacific BioLabs Report No.: 17G0155H-X03G

Implantation Test in New Zealand White Rabbits  
Pacific BioLabs Report No.: 17G0155H-X04G

The test article met the requirements for Plastic Class VI – 50°C testing as per USP <88> guidelines.

Multiple Study Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist

08/25/17

X

I approve the content of this document.  
Signed by: Zuzana Karjala

08/25/17

X

QA Review  
Signed by: Sarah Lauder