

SPONSOR 10991



3D Systems Corporation
26600 SW Parkway Ave.
Bldg 61
Wilsonville, OR 97070
United States

CERTIFICATE
CLASS VI - 50°C PLASTIC TESTS
(USP 42 <88>)

Test Article Name:	M2S-HT250
Physical Description:	Solid
Lot Number:	HB200901A
Part Number:	Not provided by Sponsor
Other Identifier:	Not provided by Sponsor
Expiration Date:	March 23rd, 2023
Sterilization Data:	Non-Sterile
Storage Conditions:	Room Temperature
Date Received:	24 Mar 2020

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits
Pacific BioLabs Report Number: 20C0404H-X02G

Systemic Injection Test in CD-1 Mice
Pacific BioLabs Report Number: 20C0404H-X03G

Implantation Test in New Zealand White Rabbits
Pacific BioLabs Report Number: 20C0404H-X04G

The test article met the requirements for Plastic Class VI – 50°C testing as per USP <88> guidelines.

Project Coordinator: Zuzana Karjala, Ph.D, RLATg, Senior Scientist

26 May 2020

X

I approve the content of this document.
Signed by: Zuzana Karjala

26 May 2020

X

QA Review
Signed by: Erika Otahal

:lc