

SPONSOR 10991



3D Systems Corporation
26600 SW Parkway Ave.
Bldg 61
Wilsonville, OR 97070
United States

REPORT DATE

09 Oct 2019

CERTIFICATE
CLASS VI - 50°C PLASTIC TESTS
(USP 42 <88>)

Test Article Name:	M2S-HT90
Physical Description:	Solid
Lot Number:	191206
Part Number:	Not provided by Sponsor
Other Identifier:	Not provided by Sponsor
Expiration Date:	Sept. 1 st , 2022
Sterilization Data:	Non-Sterile
Storage Conditions:	Room Temperature
Date Received:	09 Sep 2019

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits
Pacific BioLabs Report Number: 19I0081H-X02G

Systemic Injection Test in CD-1 Mice
Pacific BioLabs Report Number: 19I0081H-X03G

Implantation Test in New Zealand White Rabbits
Pacific BioLabs Report Number: 19I0081H-X04G

The test article met the requirements for Plastic Class VI – 50°C testing as per USP <88> guidelines.

Project Coordinator: Zuzana Karjala, Ph.D., RLATg, Senior Scientist

11 Oct 2019

X

I approve the content of this document.
Signed by: Zuzana Karjala

10 Oct 2019

X

QA Review
Signed by: Richard Bartlett

:ahc