

SPONSOR 10991



3D Systems Corporation
26600 SW Parkway Avenue
Bldg 61
Wilsonville, OR 97070
United States

CERTIFICATE
CLASS VI - 50°C PLASTIC TESTS
(USP 42 <88>)

Test Article Name: VisiJet M2E-BK70
Physical Description: Solid
Lot Number: EK200203A
Part Number: Not provided by Sponsor
Other Identifier: Not provided by Sponsor
Expiration Date: June 3rd, 2023
Sterilization Data: None
Storage Conditions: Non-Sterile
Date Received: 05 Jun 2020

Intracutaneous (Intradermal) Reactivity Test in New Zealand White Rabbits
Pacific BioLabs Report Number: 20F0069H-X02G

Systemic Injection Test in CD-1 Mice
Pacific BioLabs Report Number: 20F0069H-X03G

Implantation Test in New Zealand White Rabbits
Pacific BioLabs Report Number: 20F0069H-X04G

The test article met the requirements for Plastic Class VI – 50°C testing as per USP <88> guidelines.

Project Coordinator: Zuzana Karjala, Ph.D, RLATg, Senior Scientist

07 Aug 2020

X

I approve the content of this document.
Signed by: Zuzana Karjala

06 Aug 2020

X

QA Review
Signed by: Kenneth Tan

:lc